



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8810			
SODIUM THIOSULFATE 0.1 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 10/20/23			
Tester ENOR			
Recorded 10/20/23			
PH OF PRODUCT/SOLUTION	7.000 PH	6.000 - 12.500	10.430
Sample ID			
Tested 10/20/23			
Tester ENOR			
Recorded 10/20/23			
Actual Normality	0.1000 EQ/L	0.0995 - 0.1005	0.1000
Sample ID			
Tested 10/20/23			
Tester ENOR			
Recorded 10/20/23			
QCSP10:STANDARDIZATION OF SODIUM THIOSUL	YES QCSP10		YES
Sample ID			
Tested 10/20/23			
Tester ENOR			
Recorded 10/20/23			
QCSP35:STANDARDIZATION OF SODIUM THIOSUL	YES QCSP35		YES
Sample ID			
Tested 10/20/23			
Tester ENOR			
Recorded 10/20/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested 10/20/23			
Tester ENOR			
Recorded 10/20/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M113,T72,M55,M91,BID08251, TID41121
Sample ID			
Tested 10/20/23			
Tester ENOR			
Recorded 10/20/23			
DATE OF PREPARATION	MM/DD/YYYY		10/20/2023
Sample ID			
Tested 10/20/23			
Tester ENOR			
Recorded 10/20/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

EXPIRATION DATE	MM/DD/YY		04/30/24
	MM/DD/YYYY		
Sample ID			
Tested	10/20/23		
Tester	ENOR		
Recorded	10/20/23		
SHELF LIFE	6 MONTHS		6 MONTHS
	TIME		
Sample ID			
Tested	10/20/23		
Tester	ENOR		
Recorded	10/20/23		
TRACEABLE TO NIST SRM 136	YES		YES
	YES		
Sample ID			
Tested	10/20/23		
Tester	ENOR		
Recorded	10/20/23		
Actual Density	1.01127	1.01080 - 1.01173	1.01119
	GM/ML		
Sample ID			
Tested	10/20/23		
Tester	ENOR		
Recorded	10/20/23		
Lot #	3101008		
Made	10/20/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.


Quality Manager