Page: 1

Date: 10/18/23 at 1:14 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged	d Product:	CY229F			
3			IND QA STANDARD	(ALL SAME LOT)	
Test			Target/UOM	Range	Result
PREPARED	TO FORMU	LATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	10/18/23			
	Tester	DTRA			
	Recorded	10/18/23			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS	
			LIQUID		
			APPEARANCE		
	Sample ID				
	Tested	10/18/23			
	Tester	DTRA			
	Recorded	10/18/23			
INSTRUME	NTS USED D		LAB EQUIPMENT		P148,VF1-Z34
PREPARATION		INSTRUMENT		,	
	Sample ID				
	Tested	10/18/23			
	Tester	DTRA			
	Recorded	10/18/23			
EXPIRATION DATE		MM/DD/YYYY		12/18/2023	
			MM/DD/YYYY		
	Sample ID				
	Tested	10/18/23			
	Tester	DTRA			
	Recorded	10/18/23			
PPM COD		,,	82.0	19.0 - 145.0	58.5
			PPM		33.3
	Sample ID				
	Tested	10/18/23			
	Tester	DTRA			
	Recorded	10/18/23			
l ot	# 31016				
Mad					
iviad	IC +0/+0	, 25			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1.