

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8654 SODIUM HYDROXIDE 1.5 NORMAL SOLUTION			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 10/18/23			
Tester MHUS			
Recorded 10/18/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 10/18/23			
Tester MHUS			
Recorded 10/18/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		10/18/2023
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 10/18/23			
Tester MHUS			
Recorded 10/18/23			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	24.0
SAMPLING DEGC	DEG C		
Sample ID			
Tested 10/18/23			
Tester MHUS			
Recorded 10/18/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		10/18/2023
	MM/DD/YYYY		
Sample ID			
Tested 10/18/23			
Tester MHUS			
Recorded 10/18/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	22.0
DEG C	DEG C		
Sample ID			
Tested 10/18/23			
Tester MHUS			
Recorded 10/18/23			
NORMALITY TO 4 DEC PLACES	1.5000	1.4991 - 1.5009	1.4997
	EQUIV/LITER		
Sample ID			
Tested 10/18/23			
Tester MHUS			
Recorded 10/18/23			
QCSP7: STANDARDIZATION OF	YES		YES
AQUEOUS BASES	METHOD		
Sample ID			
Tested 10/18/23			
Tester MHUS			
Recorded 10/18/23			


Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 84	YES	YES
	YES	
Sample ID		
Tested 10/18/23		
Tester MHUS		
Recorded 10/18/23		
APPEARANCE AND COLOR	CLEAR LIQUID	PASS
	APPEARANCE	
Sample ID		
Tested 10/18/23		
Tester MHUS		
Recorded 10/18/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M112,T-4,BID-3902,PID-M91
Sample ID		
Tested 10/18/23		
Tester MHUS		
Recorded 10/18/23		
Actual Density	1.06106	1.06075 - 1.06136
	GM/ML	1.06099
Sample ID		
Tested 10/18/23		
Tester MHUS		
Recorded 10/18/23		
EXPIRATION DATE	MM/DD/YYYY	04/30/2025
	MM/DD/YYYY	
Sample ID		
Tested 10/18/23		
Tester MHUS		
Recorded 10/18/23		
Lot # 3101203		
Made 10/18/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager