

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	AME0041			
	IPA/TO	LUENE 75:25 V/V SOLUTION		
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES		YES
FILE		YES		
Sample ID				
	10/19/23			
	RSHE			
	10/19/23			
TURNED YELLOW DURIN	IG	PASS		PASS
VERIFICATION		PASS		
Sample ID				
	10/19/23			
	RSHE			
	10/19/23			
Actual Density	10/10/20	0.80571	0.80448 - 0.80695	0.80542
Actual Density		GM/ML	0.00000	0.00042
Sample ID				
	10/19/23			
	RSHE			
	10/19/23	1 4000	1 1000 1 1010	1 1000
REFRACTIVE INDEX		1.4020 RI	1.4000 - 1.4040	1.4020
O		RI		
Sample ID				
	10/19/23			
	RSHE			
	10/19/23			
APPEARANCE AND COLO	OR	CLEAR WATER WHITE		PASS
		APPEARANCE		
Sample ID				
	10/19/23			
	RSHE			
	10/19/23			
INSTRUMENTS USED DU	IRING	LAB EQUIPMENT USED		T36,T-D1,T227,VF12-98
PREPARATION		INSTRUMENT		
Sample ID				
Tested	10/19/23			
Tester	RSHE			
Recorded	10/19/23			
EXPIRATION DATE		MM/DD/YYYY		10/30/2024
		MM/DD/YYYY		
Sample ID				
	10/19/23			
	RSHE			
	10/19/23			
SHELF LIFE OF PRODUCT		1 YEAR		1 YEAR
		TIME		
Sample ID				
	10/19/23			
	RSHE			
	10/19/23			
1 COOLACO				



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	3101124
Made	10/19/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

thand St

Quality Manager