



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 8630			
Sodium Hydroxide 0.2 Normal NIST Traceable Solution			
Test	Target/UOM	Range	Result
Actual Normality	0.2000 EQ/L	0.1995 - 0.2005	0.1998
Sample ID			
Tested 10/19/23			
Tester RIRO			
Recorded 10/19/23			
QCSP7: STANDARDIZATION OF AQUEOUS BASES	YES METHOD		YES
Sample ID			
Tested 10/19/23			
Tester RIRO			
Recorded 10/19/23			
N.I.S.T. Traceable to SRM 84	Yes YES		YES
Sample ID			
Tested 10/19/23			
Tester RIRO			
Recorded 10/19/23			
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 10/19/23			
Tester RIRO			
Recorded 10/19/23			
SAMPLE COLLECTION & HANDLING PER LAB040	YES YES		YES
Sample ID			
Tested 10/19/23			
Tester RIRO			
Recorded 10/19/23			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY MM/DD/YYYY		10/19/2023
Sample ID			
Tested 10/19/23			
Tester RIRO			
Recorded 10/19/23			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	22.0
Sample ID			
Tested 10/19/23			
Tester RIRO			
Recorded 10/19/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		10/19/2023
Sample ID			
Tested 10/19/23			
Tester RIRO			
Recorded 10/19/23			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>AMBIENT TEST TEMPERATURE</b>	20.0	15.0 - 30.0	20.0
<b>DEG C</b>	DEG C		
Sample ID			
Tested	10/19/23		
Tester	RIRO		
Recorded	10/19/23		
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested	10/19/23		
Tester	RIRO		
Recorded	10/19/23		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M112,T-246,M55,BID08446,PID M91
Sample ID			
Tested	10/19/23		
Tester	RIRO		
Recorded	10/19/23		
<b>EXPIRATION DATE</b>	MM/DD/YY		10/30/25
	MM/DD/YYYY		
Sample ID			
Tested	10/19/23		
Tester	RIRO		
Recorded	10/19/23		
<b>Actual Density</b>	1.00712	1.00588 - 1.00835	1.00717
	GM/ML		
Sample ID			
Tested	10/19/23		
Tester	RIRO		
Recorded	10/19/23		
<b>Lot #</b>	3101017		
<b>Made</b>	10/19/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager