

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 7160			
POTASSIUM HYDROXIDE 0.1 NORMAL Aqueous Solution			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 10/20/23			
Tester JENC			
Recorded 10/20/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 10/20/23			
Tester JENC			
Recorded 10/20/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		10/20/2023
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 10/20/23			
Tester JENC			
Recorded 10/20/23			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	22.0
SAMPLING DEGC	DEG C		
Sample ID			
Tested 10/20/23			
Tester JENC			
Recorded 10/20/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		10/20/2023
	MM/DD/YYYY		
Sample ID			
Tested 10/20/23			
Tester JENC			
Recorded 10/20/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	22.0
DEG C	DEG C		
Sample ID			
Tested 10/20/23			
Tester JENC			
Recorded 10/20/23			
Actual Normality	0.1000	0.0995 - 0.1005	0.1000
	EQ/L		
Sample ID			
Tested 10/20/23			
Tester JENC			
Recorded 10/20/23			
QCSP7: STANDARDIZATION OF	YES		YES
AQUEOUS BASES	METHOD		
Sample ID			
Tested 10/20/23			
Tester JENC			
Recorded 10/20/23			


Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 84 (KHP)	YES		YES
	YES		
Sample ID			
Tested	10/20/23		
Tester	JENC		
Recorded	10/20/23		
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested	10/20/23		
Tester	JENC		
Recorded	10/20/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M112,T269,M55,M91,BID02259 ,TID4110
Sample ID			
Tested	10/20/23		
Tester	JENC		
Recorded	10/20/23		
EXPIRATION DATE	MM/DD/YY		04/30/25
	MM/DD/YYYY		
Sample ID			
Tested	10/20/23		
Tester	JENC		
Recorded	10/20/23		
Actual Density	1.00337	1.00321 - 1.00353	1.00342
	GM/ML		
Sample ID			
Tested	10/20/23		
Tester	JENC		
Recorded	10/20/23		
Lot #	3101869		
Made	10/20/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager