



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 209687F			
METHYLENE BLUE SOLUTION 3.2 g/L			
Test	Target/UOM	Range	Result
VERIFICATION	PASS		PASS
	PASS		
Sample ID			
Tested	10/19/23		
Tester	MHUS		
Recorded	10/19/23		
APPEARANCE AND COLOR	BLUE LIQUID APPEARANCE		PASS
Sample ID			
Tested	10/19/23		
Tester	MHUS		
Recorded	10/19/23		
Methylene Blue CAS#	7220-79-3		7220-79-3
	7220-79-3		
Sample ID			
Tested	10/19/23		
Tester	MHUS		
Recorded	10/19/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		T-35,M55
Sample ID			
Tested	10/19/23		
Tester	MHUS		
Recorded	10/19/23		
PREPARED TO FORMULATION	YES		YES
	YES		
Sample ID			
Tested	10/19/23		
Tester	MHUS		
Recorded	10/19/23		
SHELF LIFE OF PRODUCT	2 YEARS TIME		2 YEARS
Sample ID			
Tested	10/19/23		
Tester	MHUS		
Recorded	10/19/23		
EXPIRATION DATE	MM/DD/YYYY		10/30/2025
	MM/DD/YYYY		
Sample ID			
Tested	10/19/23		
Tester	MHUS		
Recorded	10/19/23		
Lot #	3101302		
Made	10/19/23		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 10/19/23 at 3:57 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager