

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	6430			
i uokugou i iouuoti	NTTRATE	STD 1000 PPM N-NO3 (1	000 PPM NITROGEN FROM NIT	RATE)
Test		Target/UOM	Range	Result
PREPARED TO FORMUL		YES	nange	YES
FILE		YES		120
Sample ID		120		
Tested	10/24/23			
Tester	GWEB			
	10/24/23			
PPM NITROGEN FROM	10/24/23	1000.0	995.0 - 1005.0	1000.0
NITRATE(NO3)		PPM	333.0 - 1003.0	1000.0
		FFIVI		
Sample ID	10/01/00			
Tested	10/24/23			
Tester	GWEB 10/24/23			
Recorded APPEARANCE AND COL		CLEAR LIQUID		PASS
AFFEARANCE AND COL	.0n	APPEARANCE		FA33
O a marke ID		APPEARANCE		
Sample ID				
Tested	10/24/23			
Tester	GWEB			
	10/24/23			
INSTRUMENTS USED DURING		INSTRUMENT		M115,VF4-N26
PREPARATION				
Sample ID				
Tested	10/24/23			
Tester	GWEB			
Recorded	10/24/23			
EXPIRATION DATE		MM/DD/YYYY		02/28/2024
		MM/DD/YYYY		
Sample ID				
Tested	10/24/23			
Tester	GWEB			
	10/24/23			
Lot # 31024	74			
Made 10/24				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA