

Made 10/23/23

Page: 1

Date: 10/24/23 at 11:50 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	DC839F			
3	TOC/TN	STANDARD 25 PPM EACH C	OMPONENT (MAKE FRESH)	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID				
Tested	10/23/23			
Tester	DTRA			
Recorded	10/23/23			
CERTIFIED PPM TOC		25.0	23.0 - 27.0	24.9
		PPM		
Sample ID				
	10/23/23			
Tester	DTRA			
	10/23/23			
CALCULATED PPM NITROGEN		25.0	24.0 - 26.0	25.0
		PPM		
Sample ID				
	10/23/23			
	DTRA			
	10/23/23			
APPEARANCE AND COL		CLEAR LIQUID		PASS
/u : =/ u u u u u u u u u u u u u u u u u u	0	APPEARANCE		
Sample ID		,		
· ·	10/23/23			
	DTRA			
	10/23/23			
INSTRUMENTS USED DU		INSTRUMENT		VF200-4,VF100-K7,VF12-87
PREPARATION	MING	INOTITOWILITY		VI 200-7, VI 100 10, VI 12 0.
Sample ID				
· ·	10/23/23			
	DTRA			
	10/23/23			
EXPIRATION DATE	10/20/20	MM/DD/YYYY		12/29/2023
EXPIRATION DATE		MM/DD/YYYY		12/20/2020
Sampla ID		IVIIVI/DD/TTT		
Sample ID	10/04/00			
	10/24/23			
	DTRA			
Recorded	10/24/23			



Page: 2

Date: 10/24/23 at 11:50 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager