

Page: 1

Date: 10/24/23 at 9:59 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produ		OUS CHLORIDE 0.02 NORMAL	SOLUTION	
Test		Target/UOM	Range	Result
PREPARED TO FOR	RMULATION ON	YES		YES
FILE		YES		
Sample	e ID			
Tested	10/24/23			
Tester	MHUS			
Record	ed 10/24/23			
Actual Normality		0.0200	0.0180 - 0.0220	0.0194
		EQ/L		
Sample	e ID			
Tested	10/24/23			
Tester	MHUS			
Record				
Actual Density		1.00743	1.00269 - 1.01218	1.00900
		GM/ML		
Sample	e ID			
Tested	10/24/23			
Tester	MHUS			
Record	ed 10/24/23			
QCSP25:STANDAR	DIZATION	YES		YES
TITANOUS CHLORI	DE	QCSP25		
Sample	e ID			
Tested	10/24/23			
Tester	MHUS			
Record	ed 10/24/23			
APPEARANCE AND COLOR		CLEAR PURPLE LIQUID		PASS
		APPEARANCE		
Sample	e ID			
Tested	10/24/23			
Tester	MHUS			
Record				
INSTRUMENTS USE	ED DURING	INSTRUMENT		GC,VF500-X40,M55
PREPARATION				
Sample	e ID			
Tested	10/24/23			
Tester	MHUS			
Record				
EXPIRATION DATE		MM/DD/YY		11/07/23
		MM/DD/YYYY		
Sample	e ID			
Tested	10/24/23			
Tester	MHUS			
Record				
Lot # 31				
	/24/23			

Page: 2

Date: 10/24/23 at 9:59 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager