10/27/23

Made

Page: 1

Date: 10/27/23 at 3:36 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

CELL WASH STD #1, 20 PPM C Target/UOM ATION ON YES	Range	
	Range	
ATION ON YES	- i turigo	Result
		YES
YES		
10/27/23		
SHAR		
10/27/23		
OR CLEAR LIQUID		PASS
APPEARANCE		
10/27/23		
SHAR		
10/27/23		
RING INSTRUMENT		P93,GC,VF500-X43
10/27/23		
SHAR		
10/27/23		
T 6 MONTHS		6 MONTHS
TIME		
10/27/23		
SHAR		
10/27/23		
MM/DD/YYYY		04/30/2024
MM/DD/YYYY		
10/27/23		
SHAR		
_		
	SHAR 10/27/23 DR CLEAR LIQUID APPEARANCE 10/27/23 SHAR 10/27/23 RING INSTRUMENT 10/27/23 SHAR 10/27/23	SHAR 10/27/23 DR CLEAR LIQUID APPEARANCE 10/27/23 SHAR 10/27/23 IRING INSTRUMENT 10/27/23 SHAR 10/27/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Pareland St