

Page: 1

Date: 10/24/23 at 9:49 AM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO #:

Customer PO # :						
Packaged Product: 8637						
SODIUM	YDROXIDE 0.3 NORMAL					
Test	Target/UOM	Range	Result			
QCSP7: STANDARDIZATION OF	YES	<u>.</u>	YES			
AQUEOUS BASES	METHOD					
Sample ID						
Tested 10/24/23						
Tester ENOR						
Recorded 10/24/23						
PREPARED TO FORMULATION ON	YES		YES			
FILE	YES					
Sample ID						
Tested 10/24/23						
Tester ENOR						
Recorded 10/24/23						
SAMPLE COLLECTION & HANDLING	YES		YES			
PER LAB040	YES					
Sample ID						
Tested 10/24/23						
Tester ENOR						
Recorded 10/24/23						
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		10/24/2023			
PURPOSES	MM/DD/YYYY					
Sample ID						
Tested 10/24/23						
Tester ENOR						
Recorded 10/24/23						
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	21.2			
SAMPLING DEGC	DEG C					
Sample ID						
Tested 10/24/23						
Tester ENOR						
Recorded 10/24/23						
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		10/24/2023			
	MM/DD/YYYY					
Sample ID						
Tested 10/24/23						
Tester ENOR						
Recorded 10/24/23						
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	20.0			
DEG C	DEG C					
Sample ID						
Tested 10/24/23						
Tester ENOR						
Recorded 10/24/23						
Actual Normality	0.3000	0.2995 - 0.3005	0.2999			
- -	EQ/L					
Sample ID						
Tested 10/24/23						
Tester ENOR						
Tester ENOR Recorded 10/24/23						

10/24/23

Made

Page: 2

Date: 10/24/23 at 9:49 AM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO #:

N.I.S.T. Traceable to SF	RM 84 (KHP)	YES YES		YES
Sample ID				
Tested	10/24/23			
Tester	ENOR			
Recorded	10/24/23			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	10/24/23			
Tester	ENOR			
Recorded	10/24/23			
INSTRUMENTS USED D	DURING	INSTRUMENT		M112,T165
PREPARATION				
Sample ID				
Tested	10/24/23			
Tester	ENOR			
Recorded	10/24/23			
Actual Density		1.01156	1.01141 - 1.01172	1.01170
		GM/ML		
Sample ID				
Tested	10/24/23			
Tester	ENOR			
Recorded	10/24/23			
EXPIRATION DATE		MM/DD/YYYY		04/30/2025
		MM/DD/YYYY		
Sample ID				
Tested	10/24/23			
Tester	ENOR			
Recorded	10/24/23			
Lot # 31019	942			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Kichard Stol

**Quality Manager**