Dackaged Products

Page: 1

Date: 10/30/23 at 5:19 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

	CELITE	STD FOR TSS 40 MG/L SOL	JTION	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample l	ID			
Tested	10/30/23			
Tester	AROS			
Recorde	d 10/30/23			
ACTUAL PPM		40.00	39.50 - 40.50	40.13
		PPM		
Sample l	ID			
Tested	10/30/23			
Tester	AROS			
Recorde				
APPEARANCE AND COLOR		SLIGHTLY TURBID LIQUID		PASS
	-	APPEARANCE		
Sample l	ID	-		
Tested	10/30/23			
Tester	AROS			
Recorde				
INSTRUMENTS USED DURING		INSTRUMENT		M108,VF4-N29
PREPARATION	·····•			
Sample	ID			
Tested	10/30/23			
Tester	AROS			
Recorde				
EXPIRATION DATE	10/00/20	MM/DD/YY		05/30/24
		MM/DD/YYYY		00/00/24
Sample l	ın	13.1411 DD/11111		
Tested	10/30/23			
Tested	10/30/23 AROS			
Recorde				
Lot # 310				
Made 10/	30/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I