

Page: 1

Date: 11/01/23 at 12:13 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	: 1763			
	CHLORID	E STANDARD 10.0 PPM W/W	AS CL (N.I.S.T)	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID				
Tested	11/01/23			
Tester	AROS			
Recorded	11/01/23			
CHLORIDE PPM		10.0	9.0 - 11.0	9.7
		PPM		
Sample ID				
Tested	11/01/23			
Tester	AROS			
Recorded	11/01/23			
APPEARANCE AND CO		CLEAR LIQUID		PASS
A I LAWARD ARD OC	LOR	APPEARANCE		17.66
Sample ID		ALLENVINOE		
Tested	11/01/23			
Tester	AROS			
Recorded	11/01/23	INCTOLIMENT		D4.42.V/E5.1.4.C
INSTRUMENTS USED I	DURING	INSTRUMENT		P143,VF5-L16
PREPARATION				
Sample ID				
Tested	11/01/23			
Tester	AROS			
Recorded	11/01/23			
EXPIRATION DATE		MM/DD/YY		05/30/24
		MM/DD/YYYY		
Sample ID				
Tested	11/01/23			
Tester	AROS			
Recorded	11/01/23			
SHELF LIFE OF PROD	UCT	SHELF LIFE		6 MONTHS
		TIME		
Sample ID				
Tested	11/01/23			
Tester	AROS			
Recorded	11/01/23			
Lot # 3110115				
Made 11/01/23				
IVIAUE 11/0.	1/20			



Page: 2

Date: 11/01/23 at 12:13 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager