

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8667 SODIUM HYDROXIDE 1.2 NORMAL LOW CARBONATE			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 11/02/23			
Tester MHUS			
Recorded 11/02/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 11/02/23			
Tester MHUS			
Recorded 11/02/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		11/02/2023
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 11/02/23			
Tester MHUS			
Recorded 11/02/23			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	20.0
SAMPLING DEGC	DEG C		
Sample ID			
Tested 11/02/23			
Tester MHUS			
Recorded 11/02/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		11/02/2023
	MM/DD/YYYY		
Sample ID			
Tested 11/02/23			
Tester MHUS			
Recorded 11/02/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	20.0
DEG C	DEG C		
Sample ID			
Tested 11/02/23			
Tester MHUS			
Recorded 11/02/23			
Product nitrogen purged	YES		YES
	YES		
Sample ID			
Tested 11/02/23			
Tester MHUS			
Recorded 11/02/23			
QCSP7: STANDARDIZATION OF	YES		YES
AQUEOUS BASES	METHOD		
Sample ID			
Tested 11/02/23			
Tester MHUS			
Recorded 11/02/23			

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

Actual Normality	1.2000 EQ/L	1.1990 - 1.2010	1.2001
Sample ID			
Tested	11/02/23		
Tester	MHUS		
Recorded	11/02/23		
N.I.S.T. Traceable to SRM 84 (KHP)	YES YES		YES
Sample ID			
Tested	11/02/23		
Tester	MHUS		
Recorded	11/02/23		
PPM INORGANIC CARBON	25.0 PPM	0.0 - 25.0	12.2
Sample ID			
Tested	11/02/23		
Tester	MHUS		
Recorded	11/02/23		
Actual Density	1.04915 GM/ML	1.04815 - 1.05035	1.04887
Sample ID			
Tested	11/02/23		
Tester	MHUS		
Recorded	11/02/23		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	11/02/23		
Tester	MHUS		
Recorded	11/02/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M55,T8667,M91,BID3902,TID4 1123	
Sample ID			
Tested	11/02/23		
Tester	MHUS		
Recorded	11/02/23		
SHELF LIFE OF PRODUCT	1 YEAR TIME		1 YEAR
Sample ID			
Tested	11/02/23		
Tester	MHUS		
Recorded	11/02/23		
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		11/30/2024
Sample ID			
Tested	11/02/23		
Tester	MHUS		
Recorded	11/02/23		
Lot #	3103005		
Made	11/02/23		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 11/02/23 at 4:19 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager