

Page: 1

Date: 11/08/23 at 4:50 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

## **Certificate of Analysis / QC Results**

Customer PO #:

Packaged Product	: CY196F			
CYANIDE STANDARD 0.2 PPM		STANDARD 0.2 PPM		
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID				
Tested	11/08/23			
Tester	MMAT			
Recorded	11/08/23			
Calculated ppm Cyanide (CN)		0.200	0.190 - 0.210	0.204
		PPM W/V		
Sample ID				
Tested	11/08/23			
Tester	MMAT			
Recorded	11/08/23			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID		2		
Tested	11/08/23			
Tester	MMAT			
Recorded	11/08/23			
INSTRUMENTS USED		INSTRUMENT		P131,VF500-X43,P72
PREPARATION	Domina	INOTHOMENT		1 101, 11 300 140,1 72
Sample ID				
Tested	11/08/23			
Tester	MMAT			
Recorded	11/08/23	MM/DD 0000/		10/10/0000
EXPIRATION DATE		MM/DD/YYYY		12/13/2023
		MM/DD/YYYY		
Sample ID				
Tested	11/08/23			
Tester	MMAT			
Recorded	11/08/23			
SHELF LIFE OF PRODUCT		1 MONTH		1 MONTH
		TIME		
Sample ID				
Tested	11/08/23			
Tester	MMAT			
Recorded	11/08/23			
Lot # 3110	810			
Made 11/0				
IVIAUE 1170	0,20			



Page: 2

Date: 11/08/23 at 4:50 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**