

Page: 1

Date: 11/09/23 at 11:29 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	ERL2371	F'		
J		US CHLORIDE SOLUTION *MA	KE FRESH*	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	ILATION ON	YES	V	YES
FILE		YES		
Sample ID				
Tested	11/09/23			
Tester	MHUS			
Recorded	11/09/23			
NORMALITY TO 4 DEC PLACES		0.0058	0.0050 - 0.0065	0.0058
		EQUIV/LITER		
Sample ID				
Tested	11/09/23			
Tester	MHUS			
Recorded	11/09/23			
Actual Density		1.05259	1.04213 - 1.06305	1.05271
		GM/ML		
Sample ID				
Tested	11/09/23			
Tester	MHUS			
Recorded	11/09/23			
APPEARANCE AND COLOR		CLEAR PURPLE LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	11/09/23			
Tester	MHUS			
Recorded	11/09/23			
INSTRUMENTS USED D	DURING	INSTRUMENT		GC,EF4-YN
PREPARATION				
Sample ID				
Tested	11/09/23			
Tester	MHUS			
Recorded	11/09/23			
EXPIRATION DATE		MM/DD/YYYY		11/23/2023
		MM/DD/YYYY		
Sample ID				
Tested	11/09/23			
Tester	MHUS			
Recorded	11/09/23			
SHELF LIFE OF PRODU		2 WEEKS		2 WEEKS
		TIME		
Sample ID				
Tested	11/09/23			
Tester	MHUS			
Recorded	11/09/23			
Lot # 31106				
Made 11/09	9/23			



Page: 2

Date: 11/09/23 at 11:29 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager