



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: ND229F			
TOC 3000 MG/L & TIC 200 MG/L			
Test	Target/UOM	Range	Result
APPEARANCE AND COLOR			
	CLEAR WATER WHITE		PASS
	APPEARANCE		
Sample ID			
Tested	11/09/23		
Tester	AROS		
Recorded	11/09/23		
CERTIFIED TOC VALUE			
	3000	2900 - 3100	2937
	PPM TOC		
Sample ID			
Tested	11/09/23		
Tester	AROS		
Recorded	11/09/23		
CERTIFIED IC VALUE			
	200	185 - 215	199
	ACT IC		
Sample ID			
Tested	11/09/23		
Tester	AROS		
Recorded	11/09/23		
INSTRUMENTS USED DURING PREPARATION			
	INSTRUMENT		M109,M108,VF200-T18,VF1-Z2 6
Sample ID			
Tested	11/09/23		
Tester	AROS		
Recorded	11/09/23		
EXPIRATION DATE			
	MM/DD/YY		01/14/2024
	MM/DD/YYYY		
Sample ID			
Tested	11/09/23		
Tester	AROS		
Recorded	11/09/23		
Lot #	3092781		
Made	11/09/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.