

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 9758-1L			
REAGENT WATER ASTM D 1193 TYPE III SUBTYPE B			
Test	Target/UOM	Range	Result
CONDUCTIVITY	MAX 0.0555 SEIMENS US/CM		<0.0555 SEIMENS
Sample ID			
Tested	11/06/23		
Tester	TBUR		
Recorded	11/06/23		
CONDUCTIVITY	MIN. 18 MEGAOHMS US/CM		<18 MEGAOHMS
Sample ID			
Tested	11/06/23		
Tester	TBUR		
Recorded	11/06/23		
TOC UG/L	MAX 200 UG/L UG/L (PPB)		<200 UG/L
Sample ID			
Tested	11/06/23		
Tester	TBUR		
Recorded	11/06/23		
SODIUM CONTENT	MAX 10 UG/L %		<10 UG/L
Sample ID			
Tested	11/06/23		
Tester	TBUR		
Recorded	11/06/23		
CHLORIDE CONTENT	MAX 10 UG/L PPM		<10 UG/L
Sample ID			
Tested	11/06/23		
Tester	TBUR		
Recorded	11/06/23		
SILICA UG/L	MAX 500 UG/L UG/L (PPB)		<500 UG/L
Sample ID			
Tested	11/06/23		
Tester	TBUR		
Recorded	11/06/23		
HETEROTROPHIC BACTERIA COUNT	MAX 10/100 CFU		< 10/100
Sample ID			
Tested	11/06/23		
Tester	TBUR		
Recorded	11/06/23		
Lot #	3092870		
Made	09/30/23		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 11/06/23 at 12:52 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager