

Page: 1

Date: 11/07/23 at 1:46 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	1862			
	CITRIC	ACID 10% W/V AQUEOUS SO	LUTION	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	11/07/23			
Tester	MHUS			
Recorded	11/07/23			
CITRIC ACID % W/V		10.0	9.5 - 10.5	10.2
		10.0		
Sample ID				
Tested	11/07/23			
Tester	MHUS			
Recorded	11/07/23			
Actual Density	, 0., = 0	1.03922	1.03655 - 1.04189	1.04031
,		GM/ML		
Sample ID		G, <u>-</u>		
Tested	11/07/23			
Tester	MHUS			
Recorded	11/07/23			
APPEARANCE AND CO		CLEAR TO PALE YELLOW		PASS
AFFEARANCE AND CO	LON	LIQUID		1 400
		APPEARANCE		
Sample ID				
Tested	11/07/23			
Tester	MHUS			
Recorded	11/07/23			
INSTRUMENTS USED D	DURING	INSTRUMENT		M100,T-169,M55,BID04099,PID
PREPARATION				P43
Sample ID				
Tested	11/07/23			
Tester	MHUS			
Recorded	11/07/23			
EXPIRATION DATE		MM/DD/YYYY		11/30/24
Sample ID				
Tested	11/07/23			
Tester	MHUS			
Recorded	11/07/23			
Lot # 31102	202			
Made 11/07				
wade +1/0/	145			



Page: 2

Date: 11/07/23 at 1:46 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager