Page: 1

Date: 11/13/23 at 4:16 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: 0367 AMMONIA STANDARD 1 PPM AS NH3			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION			YES
TILE	YES		
Sample ID			
Tested 11/1	3/23		
Tester ARO			
Recorded 11/13			
CERTIFIED PPM NH3	1.0	0.8 - 1.2	1.1
	PPM		
Sample ID			
Tested 11/1	3/23		
Tester ARO			
Recorded 11/13			
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested 11/13	3/23		
Tester ARO			
Recorded 11/13			
NSTRUMENTS USED DURING			P390,VF2-B15,M107
PREPARATION			
Sample ID			
Tested 11/10	3/23		
Tester ARO			
Recorded 11/13			
EXPIRATION DATE	MM/DD/YYYY		01/18/2024
	MM/DD/YYYY		2 2 2 2
Sample ID			
Tested 11/13	3/23		
Tester ARO			
Recorded 11/13			
	·· -		
Lot # 3111017 Made 11/13/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I