11/17/23

Made

Page: 1

Date: 11/17/23 at 3:38 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produ	ct: 6494			
	NITROGE	EN STD 100 PPM (1ML = 0.1	1MG N) SOLUTION	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample	ID			
Tested	11/17/23			
Tester	AROS			
Recorde	d 11/17/23			
CERTIFIED PPM NITROGEN:		100.0	96.0 - 104.0	98.5
		PPM		
Sample	ID			
Tested	11/17/23			
Tester	AROS			
Recorde				
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample	ID			
Tested	11/17/23			
Tester	AROS			
Recorde				
NSTRUMENTS USED DURING		INSTRUMENT		M108,GC,VF1-Z40
REPARATION				, ,
Sample	ID			
Tested	11/17/23			
Tester	AROS			
Recorde				
EXPIRATION DATE		MM/DD/YYYY		02/28/2024
		MM/DD/YYYY		
Sample	ID			
Tested	11/17/23			
Tester	AROS			
Recorde				
	1742			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1.