



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8975			
SULFIDE STANDARD 1000 PPM W/W (1ML = 1MG S2-)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 11/15/23			
Tester MHUS			
Recorded 11/15/23			
PPM SULFIDE TO 1 DECIMAL PLACE	1000.0	990.0 - 1010.0	991.8
Sample ID	PPM		
Tested 11/15/23			
Tester MHUS			
Recorded 11/15/23			
QCSP56:STANDARDIZATION SODIUM SULFIDE SO	YES		YES
Sample ID	QCSP56		
Tested 11/15/23			
Tester MHUS			
Recorded 11/15/23			
Actual Density	1.00642	1.00516 - 1.00769	1.00676
Sample ID	GM/ML		
Tested 11/15/23			
Tester MHUS			
Recorded 11/15/23			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 11/15/23			
Tester MHUS			
Recorded 11/15/23			
INSTRUMENTS USED DURING PREPARATION	LAB EQUIPMENT INSTRUMENT		M100,T-222,M55,BID3902
Sample ID			
Tested 11/15/23			
Tester MHUS			
Recorded 11/15/23			
SHELF LIFE OF PRODUCT	2 MONTHS TIME		2 MONTHS
Sample ID			
Tested 11/15/23			
Tester MHUS			
Recorded 11/15/23			
PREPARED TO STANDARD METHODS 4500-S2	YES		YES
Sample ID			
Tested 11/15/23			
Tester MHUS			
Recorded 11/15/23			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

EXPIRATION DATE	MM/DD/YY	01/20/24
	MM/DD/YYYY	
Sample ID		
Tested	11/15/23	
Tester	MHUS	
Recorded	11/15/23	
Lot #	3111301	
Made	11/15/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager