

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO # :

	DC844F			
Packaged Product:				
	TOC 2500	PPM TOTAL NITROGEN	1000 PPM SOLUTION	
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	LATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	11/16/23			
Tester	AROS			
Recorded	11/16/23			
CERTIFIED TOTAL ORG	ANIC	2500.0	2495.0 - 2505.0	2499.8
CARBON (PPM)		PPM		
Sample ID				
Tested	11/16/23			
Tester	AROS			
Recorded	11/16/23			
CERTIFIED PPM NITROGEN:		1000.0	995.0 - 1005.0	1000.0
		PPM		
Sample ID				
Tested	11/16/23			
Tester	AROS			
Recorded	11/16/23			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	11/16/23			
Tester	AROS			
	11/16/23			
INSTRUMENTS USED D	URING	INSTRUMENT		M108,VF4-N24
PREPARATION				·
Sample ID				
Tested	11/16/23			
Tester	AROS			
	11/16/23			
Lot # 31116	75			
Made 11/16,				
induc in				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA