

Made 11/16/23

Page: 1

Date: 11/16/23 at 4:34 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

D I I	Dl	VAL022F			
Packaged	Product:		GENERAL 1 000 DDM /	1 M	
		CYANIDE	STANDARD 1,000 PPM (1ML=1MG CN-) CERTIFIED	
Test			Target/UOM	Range	Result
PREPARED	TO FORMUL	ATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	11/16/23			
	Tester	JENC			
		11/16/23			
Actual ppm (Cyanide		1000.00	995.00 - 1005.00	997.43
			PPM		
	Sample ID				
	Tested	11/16/23			
	Tester	JENC			
	Recorded	11/16/23			
APPEARANCE AND COLOR		OR.	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	11/16/23			
	Tester	JENC			
	Recorded	11/16/23			
INSTRUMENTS USED DURING		JRING	INSTRUMENT		BID02259,PIDP-43,TID41100
PREPARATION	NC				
	Sample ID				
	Tested	11/16/23			
	Tester	JENC			
	Recorded	11/16/23			
EXPIRATION	I DATE		MM/DD/YY		01/21/24
			MM/DD/YYYY		
	Sample ID				
	Tested	11/16/23			
	Tester	JENC			
		11/16/23			
SHELF LIFE	OF PRODUC	СТ	SHELF LIFE		2 MONTHS
			TIME		
	Sample ID				
	Tested	11/16/23			
	Tester	JENC			
	Recorded	11/16/23			
Lot :	# 311132	23			



Page: 2

Date: 11/16/23 at 4:34 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager