

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product	DC840F				
	TOC STANDARD (KHP) 1000 PPM NON-PRESERVED (SPECIAL)				
Test		Target/UOM	Range	Result	
PREPARED TO FORMU	ILATION ON	YES		YES	
FILE		YES			
Sample ID					
Tested	11/21/23				
Tester	DTRA				
Recorded	11/21/23				
CERTIFIED MG/L TOC		1000	980 - 1020	1016	
		MG/L			
Sample ID					
Tested	11/21/23				
Tester	DTRA				
Recorded	11/21/23				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS	
		APPEARANCE			
Sample ID					
Tested	11/21/23				
Tester	DTRA				
Recorded	11/21/23				
INSTRUMENTS USED DURING		INSTRUMENT		M108,VF4-N32,M109	
PREPARATION					
Sample ID					
Tested	11/21/23				
Tester	DTRA				
Recorded	11/21/23				
EXPIRATION DATE		MM/DD/YYYY		01/26/2024	
		MM/DD/YYYY			
Sample ID					
Tested	11/21/23				
Tester	DTRA				
Recorded	11/21/23				
Lot # 31117	7109				
Made 11/21	/23				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA