

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packageo	Product:	3859			
		FLUORID	E STANDARD 2.0 PPM		
Test			Target/UOM	Range	Result
PREPARED	TO FORMU	LATION ON	YES	~	YES
FILE			YES		
	Sample ID				
	Tested	11/21/23			
	Tester	JENC			
	Recorded	11/21/23			
APPEARANCE AND COLOR		OR	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	11/21/23			
	Tester	JENC			
	Recorded	11/21/23			
INSTRUMENTS USED DURING			INSTRUMENT		P56,M118,BOTTLE,VF2-B17
PREPARATION					
	Sample ID				
	Tested	11/21/23			
	Tester	JENC			
	Recorded	11/21/23			
PPM FLUORIDE			2.00	1.80 - 2.20	2.02
			PPM		
	Sample ID				
	Tested	11/21/23			
	Tester	JENC			
	Recorded	11/21/23			
EXPIRATION DATE			MM/DD/YYYY		05/30/2024
			MM/DD/YYYY		
	Sample ID				
	Tested	11/21/23			
	Tester	JENC			
	Recorded	11/21/23			
Lot	# 31117	75			
Mad	e 11/21	/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA