



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: VUL441F			
RGNT 3 POTASSIUM IODIDE SOLUTION			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 11/20/23			
Tester MHAY			
Recorded 11/20/23			
CALCULATED GPL POTASSIUM IODIDE	20.00	19.95 - 20.05	20.00
Sample ID	GPL		
Tested 11/20/23			
Tester MHAY			
Recorded 11/20/23			
Actual Density	1.01284	1.00763 - 1.01804	1.01288
Sample ID	GM/ML		
Tested 11/20/23			
Tester MHAY			
Recorded 11/20/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested 11/20/23			
Tester MHAY			
Recorded 11/20/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		VF4-N31,M55
Sample ID			
Tested 11/20/23			
Tester MHAY			
Recorded 11/20/23			
SHELF LIFE OF PRODUCT	5 WEEKS TIME		5 WEEKS
Sample ID			
Tested 11/20/23			
Tester MHAY			
Recorded 11/20/23			
EXPIRATION DATE	MM/DD/YYYY		12/25/2023
Sample ID	MM/DD/YYYY		
Tested 11/20/23			
Tester MHAY			
Recorded 11/20/23			
Lot #	3111738		
Made	11/20/23		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 11/20/23 at 3:25 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager