

Page: 1

Date: 11/30/23 at 2:01 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product		CTANDADD 100 DDM AC NI	13 CEDUTETED COLUMION	
Test	AMMONIA	Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	11/30/23			
Tester	AROS			
Recorded	11/30/23			
CERTIFIED PPM NH3	11/30/23	100.0	96.0 - 104.0	103.5
CERTIFIED FFW NH3		PPM	90.0 - 104.0	103.3
0 1 10		PPIVI		
Sample ID				
Tested	11/30/23			
Tester	AROS			
Recorded	11/30/23			
APPEARANCE AND COLOR		CLEAR LIQUID APPEARANCE		PASS
Sample ID)			
Tested	11/30/23			
Tester	AROS			
Recorded	11/30/23			
INSTRUMENTS USED DURING		INSTRUMENT		M108,GC,VF6-29
PREPARATION				
Sample ID)			
Tested	11/30/23			
Tester	AROS			
Recorded	11/30/23			
EXPIRATION DATE		MM/DD/YYYY		03/30/2024
		MM/DD/YYYY		
Sample ID				
Tested	11/30/23			
Tester	AROS			
Recorded	11/30/23			
SHELF LIFE OF PROD		3 MONTHS		3 MONTHS
OHEEF EIFE OF TROD	001	TIME		OWORTHO
Sample ID)			
Tested	11/30/23			
Tester	AROS			
Recorded	11/30/23			
Lot # 3113	043			
Made 11/3	0/23			



Page: 2

Date: 11/30/23 at 2:01 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager