



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> TOT032F			
Mixed Anion Standard 10.0 ppm each			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 12/04/23			
Tester MMAT			
Recorded 12/04/23			
<b>PPM LITHIUM 2 DECIMAL PLACE</b>	10.00	9.95 - 10.05	10.00
Sample ID			
Tested 12/04/23			
Tester MMAT			
Recorded 12/04/23			
<b>PPM NA 2 DECIMAL PLACES</b>	10.00	9.95 - 10.05	10.00
Sample ID			
Tested 12/04/23			
Tester MMAT			
Recorded 12/04/23			
<b>PPM NH4 2 DECIMAL PLACE</b>	10.00	9.95 - 10.05	10.00
Sample ID			
Tested 12/04/23			
Tester MMAT			
Recorded 12/04/23			
<b>PPM MAGNESIUM</b>	10.00 PPM	9.95 - 10.05	10.00
Sample ID			
Tested 12/04/23			
Tester MMAT			
Recorded 12/04/23			
<b>PPM CA 2 DECIMAL PLACE</b>	10.00	9.95 - 10.05	10.00
Sample ID			
Tested 12/04/23			
Tester MMAT			
Recorded 12/04/23			
<b>PPM K 2 DECIMAL PLACE</b>	10.00	9.95 - 10.05	10.00
Sample ID			
Tested 12/04/23			
Tester MMAT			
Recorded 12/04/23			
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested 12/04/23			
Tester MMAT			
Recorded 12/04/23			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		P160,P390,VF200-4,P157,P156 ,P159
Sample ID			
Tested 12/04/23			
Tester MMAT			
Recorded 12/04/23			



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>EXPIRATION DATE</b>	MM/DD/YYYY	12/30/2024
	MM/DD/YYYY	
Sample ID		
Tested	12/04/23	
Tester	MMAT	
Recorded	12/04/23	
<b>Lot #</b>	3120451	
<b>Made</b>	12/04/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**