

Page: 1

Date: 12/08/23 at 11:14 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Package	d Product:	TEN222F			
		BROMINE	INDEX STANDARD 18 - 2	0 MG/100G (1 LITER FORMULAT	TION)
Test			Target/UOM	Range	Result
	TO FORMU	ATION ON	YES	runge	YES
FILE			YES		. =-
	Sample ID				
	Tested	12/08/23			
	Tester	AROS			
	Recorded	12/08/23			
BROMINE INDEX VALUE			19.00	15.00 - 24.00	17.64
		_	BROMINE IN	10.00 =00	.,,,,
	Sample ID		5. (S.III.) 12 II.		
	Tested	12/08/23			
	Tester	AROS			
	Recorded	12/08/23			
ΔΡΡΕΔΒΔΙ	NCE AND COL		CLEAR WATER WHITE		PASS
			APPEARANCE		
	Sample ID				
	Tested	12/08/23			
	Tester	AROS			
	Recorded	12/08/23			
INSTRUME	NTS USED D		LAB EQUIPMENT USED		M100,M108,BOTTLE
PREPARATION		J	INSTRUMENT		
	Sample ID				
	Tested	12/08/23			
	Tester	AROS			
	Recorded	12/08/23			
EXPIRATION			MM/DD/YY		03/08/24
			MM/DD/YYYY		
	Sample ID				
	Tested	12/08/23			
	Tester	AROS			
	Recorded	12/08/23			
SHELF LIFE OF PRODUCT		СТ	3 MONTHS		3 MONTHS
			TIME		
	Sample ID				
	Tested	12/08/23			
	Tester	AROS			
	Recorded	12/08/23			
Lot # 3120743					
Made 12/08/23					
ivia	,	, -			



Page: 2

Date: 12/08/23 at 11:14 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager