



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 3650			
EDTA TITRANT 0.0575 MOLAR (0.115N)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 12/04/23			
Tester ENOR			
Recorded 12/04/23			
Actual Molarity	0.0575 MOL/L	0.0570 - 0.0580	0.0574
Sample ID			
Tested 12/04/23			
Tester ENOR			
Recorded 12/04/23			
QCSP4:STANDARDIZATION EDTA SOLUTIONS	YES QCSP4		YES
Sample ID			
Tested 12/04/23			
Tester ENOR			
Recorded 12/04/23			
N.I.S.T. Traceable to SRM 915	YES YES		YES
Sample ID			
Tested 12/04/23			
Tester ENOR			
Recorded 12/04/23			
Actual Density	1.00888 GM/ML	1.00837 - 1.00939	1.00901
Sample ID			
Tested 12/04/23			
Tester ENOR			
Recorded 12/04/23			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 12/04/23			
Tester ENOR			
Recorded 12/04/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M87,T-72,BID08251,PIDM91
Sample ID			
Tested 12/04/23			
Tester ENOR			
Recorded 12/04/23			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		12/30/25
Sample ID			
Tested 12/04/23			
Tester ENOR			
Recorded 12/04/23			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT

2 YEARS
 TIME

2 YEARS

Sample ID
 Tested 12/04/23
 Tester ENOR
 Recorded 12/04/23

Lot # 3112258
Made 12/04/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager