

Page: 1

Date: 12/05/23 at 1:32 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: 7360			
Test	SSIUM IODIDE 2% W/V SOLUT: Target/UOM	ION Range	Result
PREPARED TO FORMULATION OF	YES	nange	YES
FILE	YES		120
Sample ID	120		
-			
,			
Tester BPHI			
Recorded 12/05/23	0.00	1.00 0.00	0.00
CALCULATED W/V% POTASSIUM	2.00	1.98 - 2.02	2.00
IODIDE	W/V%		
Sample ID			
Tested 12/05/23			
Tester BPHI			
Recorded 12/05/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 12/05/23			
Tester BPHI			
Recorded 12/05/23			
INSTRUMENTS USED DURING	INSTRUMENT		M87,M112,T-9
PREPARATION			,,
Sample ID			
Tested 12/05/23			
Tester BPHI			
Recorded 12/05/23			
	1.01325	1.01215 - 1.01435	1.01345
Actual Density	GM/ML	1.01215 - 1.01455	1.01545
0 1 10	GIVI/IVIL		
Sample ID			
Tested 12/05/23			
Tester BPHI			
Recorded 12/05/23			10/00/04
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		12/30/24
Sample ID			
Tested 12/05/23			
Tester BPHI			
Recorded 12/05/23			
Lot # 3112830			
Made 12/05/23			
Made 12/03/23			



Page: 2

Date: 12/05/23 at 1:32 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager