



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8384			
SILVER NITRATE 0.00282 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 12/06/23			
Tester MHUS			
Recorded 12/06/23			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 12/06/23			
Tester MHUS			
Recorded 12/06/23			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		12/06/2023
Sample ID			
Tested 12/06/23			
Tester MHUS			
Recorded 12/06/23			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	20.0
Sample ID			
Tested 12/06/23			
Tester MHUS			
Recorded 12/06/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		12/06/2023
Sample ID			
Tested 12/06/23			
Tester MHUS			
Recorded 12/06/23			
AMBIENT TEST TEMPERATURE DEGC	20.0	15.0 - 30.0	20.0
Sample ID			
Tested 12/06/23			
Tester MHUS			
Recorded 12/06/23			
ACTUAL NORMALITY TO 5 PLACES	0.00282	0.00277 - 0.00287	0.00282
Sample ID			
Tested 12/06/23			
Tester MHUS			
Recorded 12/06/23			
QCSP54:STANDARDIZATION OF LOW NORMAL SN	YES		YES
Sample ID			
Tested 12/06/23			
Tester MHUS			
Recorded 12/06/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

N.I.S.T. Traceable to SRM 919 NaCl	YES YES	YES
Sample ID		
Tested	12/06/23	
Tester	MHUS	
Recorded	12/06/23	
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE	PASS
Sample ID		
Tested	12/06/23	
Tester	MHUS	
Recorded	12/06/23	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	GC,VF3-J17,M55,
Sample ID		
Tested	12/06/23	
Tester	MHUS	
Recorded	12/06/23	
SHELF LIFE OF PRODUCT	18 MONTHS TIME	18 MONTHS
Sample ID		
Tested	12/06/23	
Tester	MHUS	
Recorded	12/06/23	
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY	06/30/2025
Sample ID		
Tested	12/06/23	
Tester	MHUS	
Recorded	12/06/23	
Lot #	3120475	
Made	12/06/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager