

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8450 SILVER NITRATE 0.282 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON YES YES			
FILE YES			
Sample ID Tested 12/06/23 Tester MHUS Recorded 12/06/23			
Actual Normality 0.2820 0.2815 - 0.2825 0.2821			
EQ/L			
Sample ID Tested 12/06/23 Tester MHUS Recorded 12/06/23			
QCSP1:STADARDIZATION OF YES YES			
SILVER NITRATE QCSP1			
Sample ID Tested 12/06/23 Tester MHUS Recorded 12/06/23			
N.I.S.T. Traceable to SRM 919 NaCl YES YES			
YES			
Sample ID Tested 12/06/23 Tester MHUS Recorded 12/06/23			
APPEARANCE AND COLOR CLEAR WATER WHITE PASS			
LIQUID			
APPEARANCE			
Sample ID Tested 12/06/23 Tester MHUS Recorded 12/06/23			
INSTRUMENTS USED DURING INSTRUMENT M113,T-58,BID-3902,PID-M91			
PREPARATION			
Sample ID Tested 12/06/23 Tester MHUS Recorded 12/06/23			
EXPIRATION DATE MM/DD/YYYY 12/30/2025			
MM/DD/YYYY			
Sample ID Tested 12/06/23 Tester MHUS Recorded 12/06/23			
Lot # 3120137			
Made 12/06/23			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 12/06/23 at 11:39 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager