Page: 1

Date: 12/05/23 at 3:10 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packag	jed Product: U	CT126F		
	C	ELITE STD FOR TSS 40 MG/L SO	DLUTION	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON			-	YES
FILE		YES		
	Sample ID			
	Tested 12/0	5/23		
	Tester ARO	S		
	Recorded 12/05	5/23		
ACTUAL	PPM	40.00	39.50 - 40.50	40.13
		PPM		
	Sample ID			
	Tested 12/0	5/23		
	Tester ARO			
	Recorded 12/0			
APPEARANCE AND COLOR		SLIGHTLY TURBID LIQUID		PASS
		APPEARANCE		
	Sample ID	, u . =, u u u u u =		
	Tested 12/0	5/23		
	Tester ARO			
	Recorded 12/0			
INSTRUMENTS USED DURING				M108,VF4-N26
PREPAR		into monitori		W1100, V1 4 1420
FILLEAN				
	Sample ID Tested 12/09	-100		
	Tested 12/09 Tester ARO			
EVDIDAT	Recorded 12/09	MM/DD/YY		06/30/24
EXPIRAT	ION DATE			06/30/24
		MM/DD/YYYY		
	Sample ID			
	Tested 12/0			
	Tester ARO			
	Recorded 12/0	5/23		
L	ot # 31205107			
	ade 12/05/23			
	440			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I