

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8430			
SILVER NITRATE 0.1 NORMAL NIST TRACEABLE SOLUTION			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 12/14/23			
Tester RIRO			
Recorded 12/14/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 12/14/23			
Tester RIRO			
Recorded 12/14/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		12/14/2023
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 12/14/23			
Tester RIRO			
Recorded 12/14/23			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	20.0
SAMPLING DEGC	DEG C		
Sample ID			
Tested 12/14/23			
Tester RIRO			
Recorded 12/14/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		12/14/2023
	MM/DD/YYYY		
Sample ID			
Tested 12/14/23			
Tester RIRO			
Recorded 12/14/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	20.0
DEG C	DEG C		
Sample ID			
Tested 12/14/23			
Tester RIRO			
Recorded 12/14/23			
Actual Normality	0.1000	0.0995 - 0.1005	0.1004
	EQ/L		
Sample ID			
Tested 12/14/23			
Tester RIRO			
Recorded 12/14/23			
QCSP53: SILVER NITRATE STD ON	YES		YES
AUTO-TITRA	METHOD		
Sample ID			
Tested 12/14/23			
Tester RIRO			
Recorded 12/14/23			

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Actual Density	1.01229 GM/ML	1.00868 - 1.01590	1.01249
Sample ID			
Tested	12/14/23		
Tester	RIRO		
Recorded	12/14/23		
VERIFICATION BY QCSP78	PASS PASS		PASS
Sample ID			
Tested	12/14/23		
Tester	RIRO		
Recorded	12/14/23		
N.I.S.T. Traceable to SRM 919 NaCl	YES YES		YES
Sample ID			
Tested	12/14/23		
Tester	RIRO		
Recorded	12/14/23		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	12/14/23		
Tester	RIRO		
Recorded	12/14/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M87,T261
Sample ID			
Tested	12/14/23		
Tester	RIRO		
Recorded	12/14/23		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		12/30/25
Sample ID			
Tested	12/14/23		
Tester	RIRO		
Recorded	12/14/23		
NORMALITY USING AUTO-TITRATOR	0.1000 EQUIV/LITER	0.0985 - 0.1015	0.1004
Sample ID			
Tested	12/14/23		
Tester	RIRO		
Recorded	12/14/23		
SHELF LIFE OF PRODUCT	SHELF LIFE TIME		2 YEARS
Sample ID			
Tested	12/14/23		
Tester	RIRO		
Recorded	12/14/23		

Lot # 3120522

Made 12/14/23



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 12/14/23 at 2:46 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager