Page: 1

Date: 12/11/23 at 11:39 AM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Pro	oduct: LY507F			
	Water 5	% v/v in Acetic Acid		
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sam	ple ID			
Test	ed 12/08/23			
Test	er MMAT			
Reco	orded 12/08/23			
H2O v/v%		5.00	4.95 - 5.05	5.00
		5.00		
Sam	ple ID			
Test	ed 12/11/23			
Test	er MMAT			
Reco	orded 12/11/23			
APPEARANCE AND COLOR		PASS		PASS
		APPEARANCE		
Sam	ple ID			
Test				
Test	er MMAT			
Reco	orded 12/11/23			
INSTRUMENTS USED DURING		INSTRUMENT		P43,VF500-X40
PREPARATION				
Sam	ple ID			
Test	•			
Test	er MMAT			
Reco	orded 12/11/23			
EXPIRATION DAT	TE	MM/DD/YYYY		12/30/2024
		MM/DD/YYYY		
Sam	ple ID			
Test	•			
Test	,,			
	orded 12/11/23			
Lot # 3	3120820			
	12/11/23			
waue -	,,			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I