

Page: 1

Date: 12/13/23 at 4:45 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	: ODP053F			
		SULFATE 0.02 NORMAL		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	12/13/23			
Tester	MMAT			
Recorded	12/13/23			
APPEARANCE AND COLOR		CLEAR TO PALE		PASS
		BLUE-GREEN LIQUID		
		APPEARANCE		
Sample ID				
Tested	12/13/23			
Tester	MMAT			
Recorded	12/13/23			
INSTRUMENTS USED I	DURING	INSTRUMENT		M115,GC,VF12-90
PREPARATION				
Sample ID				
Tested	12/13/23			
Tester	MMAT			
Recorded	12/13/23			
Actual Density		1.00367	1.00351 - 1.00384	1.00379
,		GM/ML		
Sample ID				
Tested	12/13/23			
Tester	MMAT			
Recorded	12/13/23			
CALCULATED NORMALITY		0.0200	0.0195 - 0.0205	0.0200
		EQUIV/LITER		5.52
Sample ID				
Tested	12/13/23			
Tester	MMAT			
Recorded	12/13/23			
EXPIRATION DATE	12, 13,20	MM/DD/YYYY		06/30/2024
EXI IIIAIION BAIL		MM/DD/YYYY		00/00/2021
Sample ID				
Tested	12/13/23			
Tester	MMAT			
Recorded	12/13/23			
Lot # 3120	339			
Made 12/13				
IVIGGC ==/=	- , -			



Page: 2

Date: 12/13/23 at 4:45 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager