

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:				
	CYANIDE	STANDARD 100.0 PPM (N	4AKE FRESH)	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	12/13/23			
Tester	RIRO			
Recorded	12/13/23			
APPEARANCE AND COI		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	10/10/00			
	12/13/23			
Tester	RIRO			
	12/13/23			
INSTRUMENTS USED D	URING	INSTRUMENT		BID-08446,PI-124
PREPARATION				
Sample ID				
Tested	12/13/23			
Tester	RIRO			
Recorded	12/13/23			
Actual ppm Cyanide		100.00	98.00 - 102.00	99.22
		PPM		
Sample ID				
Tested	12/13/23			
Tester	RIRO			
Recorded	12/13/23			
QCSP47:STANDARDIZATION		YES		YES
VARIOUS CYANIDE		QCSP47		
Sample ID				
Tested	12/13/23			
Tester	RIRO			
Recorded	12/13/23			
EXPIRATION DATE		MM/DD/YY		01/18/24
		MM/DD/YYYY		
Sample ID				
Tested	12/13/23			
Tester	RIRO			
Recorded SHELF LIFE OF PRODU	12/13/23	SHELF LIFE		1 MONTH
SHELF LIFE OF PRODU				
		TIME		
Sample ID	10/10/65			
Tested	12/13/23			
Tester	RIRO			
Recorded	12/13/23			
Lot # 31211	62			
Made 12/13	/23			



Page: 2 Date: 12/13/23 at 4:47 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

11hard SE

Quality Manager