Page: 1

Date: 12/11/23 at 11:38 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	8168			
SILICA STANDARD 100 PPB CERTIFIED				
Test	Tar	get/UOM	Range	Result
PREPARED TO FORMULA		YES		YES
FILE		YES		
Sample ID				
Tested 1	12/11/23			
Tester E	3PHI			
Recorded 1	12/11/23			
APPEARANCE	CLEAR	WATER WHITE		PASS
Sample ID				
Tested 1	12/11/23			
Tester E	3PHI			
Recorded 1	12/11/23			
PPB SPEC NO DEC PLAC	ES	100	90 - 110	106
		PPB		
Sample ID				
Tested 1	12/11/23			
Tester E	3PHI			
Recorded 1	12/11/23			
EXPIRATION DATE	MN	//DD/YYYY		12/30/2024
	MN	//DD/YYYY		
Sample ID				
Tested 1	12/11/23			
Tester E	3PHI			
Recorded 1	12/11/23			
Lot # 3121190				
Made 12/11/				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager