

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: SPE509F | | | |
|---|-----------------------------|-----------------|--------------|
| TOTAL SUSPENDED SOLIDS STANDARD 1000 PPM | | | |
| Test | Target/UOM | Range | Result |
| PREPARED TO FORMULATION ON | YES | | YES |
| FILE | YES | | |
| Sample ID | | | |
| Tested 12/12/23 | | | |
| Tester AROS | | | |
| Recorded 12/12/23 | | | |
| APPEARANCE AND COLOR | TURBID LIQUID APPEARANCE | | PASS |
| Sample ID | | | |
| Tested 12/12/23 | | | |
| Tester AROS | | | |
| Recorded 12/12/23 | | | |
| CALCULATED MG/L (PPM) TSS | 1000.0 MG/L | 995.0 - 1005.0 | 999.9 |
| Sample ID | | | |
| Tested 12/12/23 | | | |
| Tester AROS | | | |
| Recorded 12/12/23 | | | |
| ACT GMS CELITE | 1.0000 GM | 0.9990 - 1.0010 | 0.9999 |
| Sample ID | | | |
| Tested 12/12/23 | | | |
| Tester AROS | | | |
| Recorded 12/12/23 | | | |
| QCSP11:STANDARDIZATION OF BROMIDE-BROMAT | YES QCSP11 | | YES |
| Sample ID | | | |
| Tested 12/12/23 | | | |
| Tester AROS | | | |
| Recorded 12/12/23 | | | |
| MLS OF WATER | 1000 ML | 995 - 1005 | 1000 |
| Sample ID | | | |
| Tested 12/12/23 | | | |
| Tester AROS | | | |
| Recorded 12/12/23 | | | |
| INSTRUMENTS USED DURING PREPARATION | INSTRUMENT | | M108,VF1-Z26 |
| Sample ID | | | |
| Tested 12/12/23 | | | |
| Tester AROS | | | |
| Recorded 12/12/23 | | | |
| WEIGHT NIST TRACEABLE | PASS YES | | PASS |
| Sample ID | | | |
| Tested 12/12/23 | | | |
| Tester AROS | | | |
| Recorded 12/12/23 | | | |



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 12/12/23 at 3:02 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT

SHELF LIFE
TIME

1 YEAR

Sample ID
Tested 12/12/23
Tester AROS
Recorded 12/12/23

EXPIRATION DATE

MM/DD/YYYY
MM/DD/YYYY

12/30/2024

Sample ID
Tested 12/12/23
Tester AROS
Recorded 12/12/23

Lot # 3121148
Made 12/12/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager