

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO # :

Packaged	Product:	6490			
		NITROGE	N STANDARD 1000 PPM (1M	L=1MG N), CERTIFIED	
Test			Target/UOM	Range	Result
PREPARED	TO FORMUL	ATION ON	YES	<b>.</b>	YES
FILE			YES		
	Sample ID				
	Tested	12/14/23			
	Tester	AROS			
	Recorded	12/14/23			
PPM NITROGEN NO DECIAML			1000	980 - 1020	990
PLACES			PPM		
	Sample ID				
	Tested	12/14/23			
	Tester	AROS			
	Recorded	12/14/23			
APPEARANCE AND COLOR		.OR	CLEAR WATER WHITE		PASS
			LIQUID		
			APPEARANCE		
	Sample ID				
	Tested	12/14/23			
	Tester	AROS			
		12/14/23			
INSTRUMEN			INSTRUMENT		M108,GC,VF2-A14
PREPARATION					
	Sample ID				
	Tested	12/14/23			
	Tester	AROS			
	Recorded	12/14/23			
EXPIRATION DATE		12/11/20	MM/DD/YY		03/30/24
			MM/DD/YYYY		
	Sample ID				
	Tested	12/14/23			
	Tester	AROS			
		12/14/23			
Lot	<b>#</b> 312143	32			
Made		/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1