Page: 1

Date: 12/12/23 at 1:54 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc	ct: SPE123F			
	CHLORID	E STANDARD 10,000 PPM		
Test		Target/UOM	Range	Result
PREPARED TO FORM	IULATION ON	YES	-	YES
FILE		YES		
Sample I	D			
Tested	12/12/23			
Tester	JRIO			
Recorded	12/12/23			
CALCULATED PPM CHLORIDE (CL)		10000	9950 - 10050	10000
		PPM		
Sample I	D			
Tested	12/12/23			
Tester	JRIO			
Recorded				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample I	D			
Tested	12/12/23			
Tester	JRIO			
Recorded				
INSTRUMENTS USED		INSTRUMENT		M103,T25
PREPARATION				
Sample I	n			
Tested	12/12/23			
Tester	JRIO			
Recorded				
EXPIRATION DATE	12/12/20	MM/DD/YYYY		12/30/2024
EXITICATION DATE		MM/DD/YYYY		12/00/2024
Sample I	n	IVIIVII D DI T T T T		
Tested	12/12/23			
Tester	JRIO			
Recorded				
Lot # 3120591				
Made 12/1	12/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I