

Page: 1

Date: 12/13/23 at 2:08 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:				
Packaged Product: 8407				
SILVER 1	NITRATE 0.025 NORMAL			
Test	Target/UOM	Range	Result	
Actual Normality	0.0250	0.0245 - 0.0255	0.0248	
	EQ/L			
Sample ID				
Tested 12/13/23				
Tester RIRO				
Recorded 12/13/23				
QCSP53: SILVER NITRATE STD ON	YES		YES	
AUTO-TITRA	METHOD			
Sample ID				
Tested 12/13/23				
Tester RIRO				
Recorded 12/13/23				
QCSP54:STANDARDIZATION OF	YES		YES	
LOW NORMAL SN	QCSP54			
Sample ID				
Tested 12/13/23				
Tester RIRO				
Recorded 12/13/23				
QCSP78:CHLORIDE SQC METHOD	YES		YES	
TO VERIFY SN	QCSP78			
Sample ID				
Tested 12/13/23				
Tester RIRO				
Recorded 12/13/23				
N.I.S.T. Traceable to SRM 919	Yes		YES	
	YES			
Sample ID				
Tested 12/13/23				
Tester RIRO				
Recorded 12/13/23	V/50		\/F0	
PREPARED TO FORMULATION ON	YES		YES	
FILE	YES			
Sample ID				
Tested 12/13/23				
Tester RIRO				
Recorded 12/13/23	YES		YES	
SAMPLE COLLECTION & HANDLING PER LAB040	YES		169	
	IES			
Sample ID Tested 12/13/23				
Tested 12/13/23 Tester RIRO				
Recorded 12/13/23				
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		12/13/2023	
PURPOSES	MM/DD/YYYY		12, 10,2020	
Sample ID	191101/20/1111			
Tested 12/13/23				
Tester RIRO				
Recorded 12/13/23				
. 1555. 554 12, 10,20				



Page: 2

Date: 12/13/23 at 2:08 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	20.0
SAMPLING DEGC	DEG C		
Sample ID			
Tested 12/13/23			
Tester RIRO			
Recorded 12/13/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		12/13/2023
	MM/DD/YYYY		
Sample ID			
Tested 12/13/23			
Tester RIRO			
Recorded 12/13/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	20.0
DEG C	DEG C		
Sample ID			
Tested 12/13/23			
Tester RIRO			
Recorded 12/13/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
74 1 2740 4102 7412 002011	LIQUID		
	APPEARANCE		
Sample ID	74 1 274 0 4 0 2		
Tested 12/13/23			
Tester RIRO			
Recorded 12/13/23			
INSTRUMENTS USED DURING	INSTRUMENT		VF12-82,M91,M55
PREPARATION	INCTROMENT		V1 12-02,IVIO 1,IVIOO
Sample ID			
Tested 12/13/23			
Tester RIRO			
Recorded 12/13/23	4.004.04	4.00400 4.00000	4 00470
Actual Density	1.00181	1.00126 - 1.00236	1.00179
	GM/ML		
Sample ID			
Tested 12/13/23			
Tester RIRO			
Recorded 12/13/23			
Lot # 3120853			
Made 12/13/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Page: 3

Date: 12/13/23 at 2:08 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager