

Page: 1

Date: 12/19/23 at 4:20 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results Customer PO #:

oustomer i o

Packaged Product: 1761			
	DE STANDARD 0.5 PPM as C	1	
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 12/19/23			
Tester MMAT			
Recorded 12/19/23			
N.I.S.T. Traceable	YES		YES
	YES		
Sample ID			
Tested 12/19/23			
Tester MMAT			
Recorded 12/19/23			
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested 12/19/23			
Tester MMAT			
Recorded 12/19/23			
INSTRUMENTS USED DURING	LAB EQUIPMENT		P349,VF2-A4
PREPARATION	INSTRUMENT		
Sample ID			
Tested 12/19/23			
Tester MMAT			
Recorded 12/19/23			
EXPIRATION DATE	MM/DD/YY		06/30/24
	MM/DD/YYYY		
Sample ID			
Tested 12/19/23			
Tester MMAT			
Recorded 12/19/23			
SHELF LIFE OF PRODUCT	SHELF LIFE		6 MONTHS
	TIME		
Sample ID			
Tested 12/19/23			
Tester MMAT			
Recorded 12/19/23			
CALCULATED PPM CHLORIDE (CL)	0.5	0.4 - 0.6	0.5
	PP		
Sample ID			
Tested 12/19/23			
Tester MMAT			
Recorded 12/19/23			
Lot # 3121971			
Made 12/19/23			



Page: 2

Date: 12/19/23 at 4:20 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager