

Page: 1

Date: 12/19/23 at 5:48 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc	t: CY157F			
	CYANIDE	STANDARD 100 PPM		
Test		Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES		YES
FILE		YES		
Sample ID)			
Tested	12/19/23			
Tester	SHAR			
Recorded	12/19/23			
Actual PPM		100.00	98.00 - 102.00	99.46
		PPM		
Sample ID				
Tested	12/19/23			
Tester	SHAR			
Recorded	12/19/23			
QCSP47:STANDARDIZATION		YES		YES
VARIOUS CYANIDE		QCSP47		
Sample ID)			
Tested	12/19/23			
Tester	SHAR			
Recorded	12/19/23			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID)			
Tested	12/19/23			
Tester	SHAR			
Recorded	12/19/23			
INSTRUMENTS USED	DURING	INSTRUMENT		P172,BID079,TID41124
PREPARATION				
Sample ID)			
Tested	12/19/23			
Tester	SHAR			
Recorded				
SHELF LIFE OF PROD		6 WEEKS		6 WEEKS
		TIME		
Sample ID)			
Tested	12/19/23			
Tester	SHAR			
Recorded				
EXPIRATION DATE	-	MM/DD/YYYY		01/30/2024
		MM/DD/YYYY		
Sample ID)			
Tested	12/19/23			
Tester	SHAR			
Recorded				
Lot # 3121				
Made 12/1	3143			



Page: 2

Date: 12/19/23 at 5:48 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager