

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product	275-02	2F		
-		CHLORIC ACID N/10 STANDA	RDIZED SOLUTION (0.1 NORMA	AL)
Test		Target/UOM	Range	Result
PREPARED TO FORMU	ILATION	YES	-	YES
		YES		
Sample ID				
Tested	12/20/23			
Tester	MHUS			
Recorded	12/20/23			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		APPEARANCE		
Sample ID				
Tested	12/20/23			
Tester	MHUS			
Recorded	12/20/23			
INSTRUMENTS USED [DURING	INSTRUMENT		VF1-Z43
PREPARATION				
Sample ID				
Tested	12/20/23			
Tester	MHUS			
Recorded	12/20/23			
NORMALITY TO 4 DEC PLACES		0.1000	0.0995 - 0.1005	0.1004
		EQUIV/LITER		
Sample ID				
Tested	12/20/23			
Tester	MHUS			
Recorded	12/20/23			
Actual Density	12/20/20	1.00008	0.99992 - 1.00024	1.00002
riotaal Donolty		GM/ML		
Sample ID				
Tested	12/20/23			
Tester	MHUS			
Recorded	12/20/23			
EXPIRATION DATE	12/20/25	MM/DD/YYYY		12/30/2025
EAFIRATION DATE		MM/DD/YYYY		12/30/2023
Operate JD				
Sample ID	10/00/00			
Tested	12/20/23			
Tester	MHUS			
Recorded	12/20/23			
Lot # 31215				
Made 12/20)/23			



Page: 2 Date: 12/20/23 at 4:37 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1 chard SE

Quality Manager