12/20/23

Made

Page: 1

Date: 12/20/23 at 4:04 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc	ct: ERL252F			
-	NITROGE	N STOCK STANDARD 100 PI	PM W/W IN ISOOCTANE	
Test		Target/UOM	Range	Result
PREPARED TO FORM	MULATION ON	YES		YES
FILE		YES		
Sample I	D			
Tested	12/20/23			
Tester	AROS			
Recorded	d 12/20/23			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample I	D			
Tested	12/20/23			
Tester	AROS			
Recorded	d 12/20/23			
INSTRUMENTS USED DURING		INSTRUMENT		M115,M100
PREPARATION				
Sample I	D			
Tested	12/20/23			
Tester	AROS			
Recorded	d 12/20/23			
PPM NITROGEN W/W		100.00	98.00 - 102.00	99.96
		PPM		
Sample I	D			
Tested	12/20/23			
	AROS			
Tester	ARUS			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager** 

hichard Stol