Page: 1

Date: 12/29/23 at 11:35 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc	t: LAB3%PE	LAB3%PEROXIDEF			
•		KIDE FOR CLEANING PURPO	IG PURPOSES AQUA USE ONLY (1000L)		
Test		Target/UOM	Range	Result	
PREPARED TO FORMULATION ON		YES		YES	
FILE		YES			
Sample II	D				
Tested	12/29/23				
Tester	MHAY				
Recorded	12/29/23				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS	
		APPEARANCE			
Sample II	D				
Tested	12/29/23				
Tester	MHAY				
Recorded	12/29/23				
INSTRUMENTS USED DURING		INSTRUMENT		M100,M108,VF1-Z24	
PREPARATION					
Sample II	D				
Tested	12/29/23				
Tester	MHAY				
Recorded	12/29/23				
Lot # 3122	2865				
Made 12/2	29/23				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager